



---



In King County, WA

# After Use in an Emergency

1. DO replace the pads. The device will make a chirping sound until the pads are replaced.
2. DO NOT remove the battery. All data will be erased. The battery is not causing the chirping sound.
3. Contact Emergency Medical Services (EMS), Public Health - Seattle & King County **as soon as possible** at [AED@kingcounty.gov](mailto:AED@kingcounty.gov) or (206) 296-4693. Tell them your AED was used and needs a download. Leave a voice mail with your phone number if no one answers. **If you use an AED management company please contact EMS first. Contact us even if no shock was delivered.** (Required in King County, WA per RCW 70.54.310)
4. EMS staff will arrange to come to your site to download the report from your device.

Emergency Medical Services Division  
401 Fifth Avenue, Suite 1200  
Seattle, WA 98104  
206-296-4693

Public Health   
Seattle & King County

## SHOCKINGLYSIMPLE

-~♥~ RESTART A HEART ~-♥~